


## TEST REPORT

No.  
19F18

Date  
13 Dec 2019

Page n.  
1 (2)

Annexes n.  
2



### **Determination of the filtration performance of tissue element** **CODE ANTIPO 180-30**

Requested by: T.I.E. Tessitura Industriale Europea S.P.A.

**Dep. CTS Laboratori**  
tel. +39 02 66409991 fax +39 02 6194115 info@ctslab.eu

**Air Filter Testing Laboratory**  
Via Zucchi, 39/C 20095 Cusano Milanino (MI) - Italy

**Bio Analysis Laboratory**  
Via Zucchi, 19 20095 Cusano Milanino (MI) - Italy

## TEST REPORT

No.  
19F18

Date  
13 Dec 2019

Page n.  
2 (2)

Annexes n.  
2



**Requested by:** T.I.E. Tessitura Industriale Europea S.P.A.  
**Address:** Località Petrona 18  
50038 Scarperia e San Pietro (FI) - Italy

**Sample(s):** Delivered sample, No. 1 piece

### Purpose of the test

Measurement of the aerodynamic characteristics and performance of a tissue element for different particle size.

### Sample Supplied

The sample tissue is identified with the code **ANTIPO 180-30**. The frame containing the tissue element have a dimension of 595mm (width) x 595mm (height) x 10mm (depth).

### Laboratory's remarks

The customer delivered one tissue element. Sample arrived in good condition with no obvious damage.

### Test Method

The test sample consist of one tissue element mounted in the test duct inside a metal plate with dimension 610mm (width) x 610mm (height) x 75mm (thickness).

The air flow resistance, as a function of the air flow rate, was measured at **200 m<sup>3</sup>/h**.

The spectral efficiency of the filter was measured with optical particle counters (OPC) for the following particle sizes: **0.5 / 1.0 / 2.0 / 5.0 / 10.0 / 25.0 µm**. This test was performed at the air flow rate of **200 m<sup>3</sup>/h**;

The test aerosol used to determine the spectral efficiency was **KCl (Potassium chloride)**.

The test was performed in date:

**12.12.2019**


**Resistance to Airflow**

**12.12.2019**

**Fractional Efficiency by Particle Size**

Results of the test:

**See annexes: 1**



**Laboratory Tester**  
Davide Degiorgi



**Laboratory Manager**  
Luca Zucchelli



Requested by: **T.I.E. Tessitura Industriale Europea S.P.A.**  
Località Petrona 18  
50038 Scarperia e San Pietro (FI)  
Italy

Date of test:	<b>12 Dec 2019</b>		
Temperature:	<b>19.3 °C</b>	Relative humidity:	<b>46.2 %</b>
		Atm. Pressure:	<b>98,698 kPa</b>

TEST DATA DETAILS			
Type of test:	<b>Fractional Efficiency by Particle Size</b>		
Air Filter:	<b>ANTIPO 180-30</b>	Article number:	---
Dimension WxHxD (mm):	<b>595x595x10</b>	Filter state:	<b>Untreated</b>
Filter material:	---	Airflow rate (m <sup>3</sup> /h):	<b>200</b>
Test aerosol:	<b>KCl 20%</b>	Pressure drop (Pa):	<b>0</b>
Cycles No:	<b>6</b>	Sampling time:	<b>30 s</b>

TEST RESULTS					
Particle Size		Efficiency Measurement by Particle Size			
Interval (µm)	Interval (nm)	Concentration (N/cf)		Efficiency (%)	Uncertainty U
		Upstream	Downstream		
0.5	500	15296685	15141416	1.01	± 0,139
1.0	1000	13907428	13699815	1.50	± 0,332
2.0	2000	13830000	13492885	2.44	± 0,124
5.0	5000	4393983	3512893	20.05	± 0,596
10.0	10000	5768	3851	32.71	± 1,812
25.0	25000	670	287	56.99	± 0,513

*The uncertainty of the measured efficiencies is reported on a 95% confidence limit.*






**Dep. CTS Laboratory**  
**Air Filter Testing Laboratory**  
Via Zucchi, 39/C  
20095 Cusano Milanino (MI) - Italy

Test Report N. 19F18

**ANNEX 1**

Pag. 2 di 2

### TEST RESULTS



**Laboratory Tester**

Davide Degiorgi

**Laboratory Manager**

Luca Zucchelli

*The report applies to the tested device only. The performance results cannot by themselves be quantitatively applied to predict filtration performance in all real life environments. This report must not be reproduced without the written approval of the CTS Laboratory.*



*Requested by:* **T.I.E. Tessitura Industriale Europea S.P.A.**  
Località Petrona 18  
50038 Scarperia e San Pietro (FI)  
Italy

INSTRUMENT USED				
Instrument	Type Code	Serial Number	Calibration Date	Used
Optical particle counter	Lasair-210	22210-1090-40	07 Apr, 2014	---
	Lasair III 5100	65360	16 Mar, 2018	✓
	Lasair III 110	125399	06 Dec, 2017	---
Differential pressure gage	Dwyer Magnehelic	W51WRD	05 Jan, 2017	---
	KIMO AMI 300	12032877	16 Mar, 2018	✓
Hygrometer	KIMO AMI 300	12017154	16 Mar, 2018	✓
Temperature meter	KIMO AMI 300	12017154	16 Mar, 2018	✓
Air flow meter	KIMO DBM 610	12030245	16 Mar, 2018	✓
	TSI 8380	T83801319007	08 May, 2014	---
Dilution system	VKL-10 Palas	5982	03 Dec, 2014	---
	VKL-100s Palas	3772	04 Jul, 2014	---
Photometer	ATI-2i	24299	21 Aug, 2018	---
Aerosol generator	AGK 2000 Palas	8220	29 Mar, 2017	✓
Aerosol generator	AGP 2.0ip Palas	0988	* n.p.	---

\* not provided

*The report applies to the tested device only. The performance results cannot by themselves be quantitatively applied to predict filtration performance in all real life environments. This report must not be reproduced without the written approval of the CTS Laboratory.*